


Schedule of Accreditation

issued by

United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 0043 Accredited to ISO/IEC 17025:2005	Roxspur Measurement and Control Ltd	
	Issue No: 058 Issue date: 09 August 2017	
	2 Downgate Drive Sheffield South Yorkshire S4 8BT	Contact: Mr M Donnelly Tel: +44 (0)114 244 2521 Fax: +44 (0)114 243 4838 E-Mail: Mark.Donnelly@ttelectronics.com Website: www.roxspur.com

Calibration performed by the Organisations at the locations specified below

Locations covered by the organisation and their relevant activities

Laboratory locations:

Location details	Activity	Location code
Address 2 Downgate Drive Sheffield South Yorkshire S4 8BT	Local contact Mr Mark Donnelly Tel: +44 (0)114 244 2521 Fax: +44 (0)114 243 4838 Email: Mark.Donnelly@ttelectronics.com Website: www.roxspur.com	Electrical Flow Pressure Temperature P1

Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Local contact Mr Mark Donnelly Tel: +44 (0)114-244 2521 Fax: +44 (0)114-243 4838 Email: mark.donnelly@ttelectronics.com Website: www.roxspur.com	Electrical Pressure Temperature S



0043

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 058 Issue date: 09 August 2017

Calibration performed by the Organisation at the locations specified

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL				
DC Voltage	0 V to 200 mV 200 mV to 2 V 2 V to 20 V 20 V to 200 V 200 V to 1000 V	15 ppm + 0.70 μ V 10 ppm + 0.60 μ V 10 ppm 15 ppm 15 ppm		P1
DC Current	0 V to 30 V	5.0 mV		S
	0 μ A to 200 μ A 200 μ A to 2 mA 2 mA to 20 mA 20 mA to 200 mA 200 mA to 2 A	50 ppm + 2.0 nA 75 ppm + 15 nA 75 ppm 100 ppm 200 ppm		P1
	0 mA to 100 mA	0.010 mA		S
Generation	320 mA to 3.2 A 3.2 A to 10 A 10 A to 20 A	550 ppm + 150 μ A 500 ppm + 1.1 mA 0.11 % + 5.2 mA		P1
DC Resistance Measurement	20 A to 1000 A	0.31 %	For the calibration of clampmeters only	
	0 Ω to 20 Ω 20 Ω to 2 k Ω 2 k Ω to 20 k Ω 20 k Ω to 200 k Ω 200 k Ω to 2 M Ω 2 M Ω to 20 M Ω 20 M Ω to 200 M Ω 200 M Ω to 2 G Ω	100 ppm + 2.0 m Ω 20 ppm + 2.0 m Ω 20 ppm 30 ppm 50 ppm 200 ppm 700 ppm 0.10 %		P1
AC Voltage	10 mV to 200 mV 20 Hz to 10 kHz 10 kHz to 100 kHz	280 ppm + 8.0 μ V 550 ppm + 10 μ V		P1
	200 mV to 200 V 20 Hz to 30 kHz 30 kHz to 100 kHz	400 ppm 480 ppm		
	200 V to 1100 V 45 Hz to 10 kHz 10 kHz to 30 kHz	450 ppm 510 ppm		



0043

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 058 Issue date: 09 August 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
ELECTRICAL (cont'd)				
AC Current	20 Hz to 1 kHz 10 μ A to 200 μ A 200 μ A to 200 mA 200 mA to 2 A	250 ppm + 41 nA 350 ppm 600 ppm		P1
	1 kHz to 5 kHz 10 μ A to 200 μ A 200 μ A to 200 mA 200 mA to 2 A	800 ppm + 30 nA 700 ppm 0.13 %		P1
Generation	10 Hz to 3 kHz 2 A to 20 A	0.25 % + 7.0 mA		
Frequency	40 Hz to 100 Hz 20 A to 1000 A	0.50 %	For the calibration of clampmeters only	
Time interval	0.1 Hz to 120 MHz	0.7 ppm		P1
	60 s to 8 hrs 60 s to 8 hrs	0.40 s 0.60 s		P1 S
Electrical calibration of temperature indicators, controllers and recorders for the following sensors:				
Noble metal thermocouples				
Types R and S	0 °C to 200 °C 200 °C to 1700 °C	0.50 °C 0.40 °C	with cold junction compensation	P1
Type B	600 °C to 1700 °C	0.40 °C	with cold junction compensation	P1
Type R and S	0 °C to 200 °C 200 °C to 1700 °C	0.60 °C 0.60 °C	with cold junction compensation	S
Type B	600 °C to 1700 °C	1.0 °C	with cold junction compensation	S
Base metal thermocouples				
	-200 °C to 0.0 °C 0 °C to 1372 °C	0.26 °C 0.20 °C	with cold junction compensation	P1
	-200 °C to 0.0 °C 0 °C to 1372 °C	0.80 °C 0.60 °C	with cold junction compensation	S



0043

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 058 Issue date: 09 August 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
Electrical calibration of temperature indicators, controllers and recorders for the following sensors: (cont'd)				
Pt 100	-200 °C to +850 °C	0.050 °C		P1
	-200 °C to +850 °C	0.50 °C		S
FLOW				
Flow rate - Gas	0.005 l/min to 0.25 l/min 0.25 l/min to 2.5 l/min 2.5 l/min to 50 l/min	0.53 % 0.34 % 0.43 %	Calibrations medium: Compressed Air Oxygen Argon Carbon Dioxide Nitrous Oxide Helium Nitrogen	P1
PRESSURE				
Gas pressure (absolute)				
Calibration of pressure indicating instruments and gauges	3.5 kPa to 120 kPa 120 kPa to 3.1 MPa	0.015 % + 71 Pa 0.015 % + 71 Pa	Calibration of pressure devices with an electrical output may be undertaken	P1
	3 kPa to 96 kPa 96 kPa to 105 kPa 105 kPa to 2.1 MPa	3.0 kPa 400 Pa 3.0 kPa		S
Gas pressure (gauge)				
Calibration of pressure indicating instruments and gauges	-90 kPa to -3.5 kPa -3.5 kPa to -2.5 kPa -2.5 kPa to 2.5 kPa 2.5 kPa to 3.5 kPa 3.5 kPa to 2.5 MPa	0.015 % 71 Pa 5.5 Pa 71 Pa 0.015 %		P1
	- 90 kPa to + 2.0 MPa	2.5 kPa		S
Hydraulic pressure (absolute)				
Calibration of pressure indicating instruments and gauges	700 kPa to 6.1 MPa 6.1 MPa to 120.1 MPa	0.016 % + 71 Pa 0.020 % + 71 Pa		P1
	10 kPa to 60 MPa	220 kPa		S
Hydraulic pressure (gauge)				
Calibration of pressure indicating instruments and gauges	600 kPa to 6 MPa 6 MPa to 120 MPa	0.016 % 0.020 %		P1
	0 to 70 MPa	220 kPa		S



0043

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory

Issue No: 058 Issue date: 09 August 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE				
Resistance thermometers	- 196 °C 0 °C to - 80 °C Ice Point (0 °C) Triple Point of water (0.01 °C) Gallium Melt Point (29.7646 °C) 0 °C to 300 °C	0.040 °C 0.050 °C 0.031 °C 0.012 °C 0.010 °C 0.050 °C		P1
Platinum thermocouples	0 °C to 1100 °C 1100 °C to 1600 °C 1064.18 °C 1553.5 °C	0.50 °C 1.8 °C 0.50 °C 1.8 °C	Au and Pd wire bridge measurements	P1
	0 °C to 1100 °C 1100 °C to 1600 °C	2.0 °C 4.0 °C		S
	Fixed point calibrations Gallium Melt Point (29.7646 °C) Tin (231.9 °C) Zinc (419.5 °C)	0.35 °C 0.50 °C 0.50 °C		P1
Other thermocouples	- 196 °C 0 °C to - 80 °C Gallium Melt Point (29.7646 °C) 0 °C to 300 °C 300 °C to 650 °C 650 °C to 1100 °C 1100 °C to 1300 °C	0.26 °C 0.15 °C 0.10 °C 0.15 °C 0.25 °C 1.0 °C 2.5 °C		P1
	- 80 °C to + 200 °C 200 °C to 1000 °C 1000 °C to 1300 °C	1.0 °C 3.0 °C 5.0 °C		S
Compensating and extension cables for Noble metal thermocouples Base metal thermocouples	0 °C to 40 °C 0 °C to 40 °C	0.26 °C 0.15 °C		P1
Liquid-in-glass thermometers	- 80 °C to - 40 °C - 40 °C to 0 °C Ice point (0 °C) 0 °C to 100 °C 100 °C to 300 °C	0.11 °C 0.050 °C 0.050 °C 0.050 °C 0.050 °C		P1
Optical Pyrometers	800 °C to 1100 °C 1100 °C to 1600 °C	5.0 °C 8.0 °C	Using effective wavelength of 0.65 µm to 0.66 µm	P1
Radiation thermometers	500 °C to 1100 °C 1100 °C to 1600 °C	4.0 °C 5.0 °C		P1



0043

Accredited to
ISO/IEC 17025:2005

Schedule of Accreditation
issued by
United Kingdom Accreditation Service
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

Roxspur Measurement and Control Calibration Laboratory
Issue No: 058 Issue date: 09 August 2017

Calibration performed by the Organisation at the locations specified

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ($k = 2$)	Remarks	Location Code
TEMPERATURE (cont'd)				
Electronic thermometers with sensors	Range as per sensors	As for sensors		P1
Electronic thermometers with sensors	0 °C to 200 °C	0.50 °C		S
Analogue	200 °C to 600 °C	1.3 °C		
Digital	600 °C to 1100 °C	2.0 °C		
	1100 °C to 1300 °C	3.4 °C		
Metal block calibrators	- 100 °C to + 300 °C	0.050 °C		P1
	300 °C to 1100 °C	1.0 °C		
Temperature surveys				
Autoclaves, incubators and Freezers	- 80 °C to + 200 °C	1.0 °C	Single and multipoint monitoring probes. Time dependent temperature profiling.	S
Ovens, furnaces and environmental chambers	0 °C to 600 °C	1.1 °C		
	600 °C to 1100 °C	1.9 °C		
	1100 °C to 1600 °C	3.5 °C		
END				